

We are very grateful for your efforts and advice about JGS station. After the audit, we are taking actions mainly on aerosol inlets for drying and T/RH monitoring according to your recommendations.

<CPC>

TSI dilutor provides two capillaries of dilution ratio of 1/20 and 1/100. We are using 1/20 capillaries, and TSI agency said we can check the dilution ratio with digital flow meter if it passes 1/20 of the initial flowrate. And we also would like to check if the dilution ratio can be changed to 2-3 in the near future as you recommended.